

ABSTRACT

An adjustment of hearing device systems is achieved by an interactive adaptation that is undertaken during the operation. A classifier recognizes various auditory situations and starts interactive adaptation procedures, in the framework of which various settings are to be evaluated. Only settings that fit the auditory situation are thereby offered. As a result of the evaluations, an optimal hearing aid settings results for the respective auditory situation.